
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001260
15:54 16/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:17
PRESSURIZING 4:18
STERILIZING 5:38
136.0 C 221kPa 5:38
136.2 C 221kPa 6:08
136.4 C 225kPa 6:38
136.3 C 224kPa 7:08
136.3 C 224kPa 7:38
136.3 C 223kPa 8:08
136.4 C 224kPa 8:38
136.4 C 225kPa 9:08
136.4 C 225kPa 9:08
VENTING 9:08
AIR DRYING 10:05
CYCLE COMPLETE 17:27

Digital Signature #
076B773D33FC82A9
