
Midmark M11 - v1.0.0
Total Cycles: 1671

Name: RITTER 1
Sterilizer ID: V2178198
02/24/2022 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 186.1 F 3.3
4:00 202.6 F 7.6
6:00 217.5 F 12.9
8:00 227.8 F 7.2
10:00 239.4 F 12.5
12:00 237.9 F 9.7
14:00 237.4 F 8.5
16:00 249.7 F 14.8
18:00 261.9 F 21.7
20:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.5 F 28.6
2:00 271.7 F 28.3
3:00 272.8 F 29.2
4:00 272.4 F 28.6

Min 270.9 F 27.9
Max 273.3 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.1 F 0.0
5:00 228.1 F 0.0
10:00 229.6 F 0.0
15:00 227.7 F 0.0
20:00 230.3 F 0.0
25:00 231.9 F 0.0
30:00 208.6 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

02/24/2022 03:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____