

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3936

Name: RITTER 1  
Sterilizer ID: V2178198  
10/23/2023 12:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 154.7 F 0.0  
2:00 154.8 F 2.2  
4:00 166.9 F 5.7  
6:00 192.3 F 9.9  
8:00 215.5 F 5.4  
10:00 227.6 F 8.1  
12:00 239.1 F 12.9  
14:00 236.6 F 9.8  
16:00 239.6 F 9.0  
18:00 247.7 F 14.1  
20:00 259.2 F 20.4  
22:00 269.7 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 272.3 F 29.1  
2:00 271.8 F 28.5  
3:00 272.1 F 29.0  
4:00 272.2 F 28.8

Min 271.3 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 227.4 F 0.0  
10:00 236.7 F 0.0  
15:00 222.1 F 0.0  
20:00 215.9 F 0.0  
25:00 212.8 F 0.0  
30:00 209.9 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 22:51  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:01:49  
CYCLE COMPLETE

10/23/2023 01:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_