

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3417

Name: RITTER 1  
Sterilizer ID: V2178198  
06/20/2023 03:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
2:00 198.2 F 4.0  
4:00 217.0 F 9.7  
6:00 229.8 F 7.1  
8:00 239.1 F 11.9  
10:00 238.1 F 10.0  
12:00 237.7 F 8.4  
14:00 251.6 F 15.9  
16:00 263.2 F 22.8  
18:00 271.6 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.3 F 28.7  
2:00 271.7 F 28.6  
3:00 272.2 F 28.7  
4:00 271.5 F 28.4  
  
Min 271.4 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.0 F 0.0  
5:00 238.4 F 0.0  
10:00 221.6 F 0.0  
15:00 219.0 F 0.0  
20:00 217.3 F 0.0  
25:00 216.3 F 0.0  
30:00 215.4 F 0.0  
Drying Complete

FILLING: 2:40  
HEATING: 18:04  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:57:54  
CYCLE COMPLETE

06/20/2023 04:23 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_