
Midmark M11 - v1.0.0
Total Cycles: 3308

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2023 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 195.1 F 2.1
4:00 214.2 F 7.9
6:00 229.2 F 11.6
8:00 235.8 F 10.4
10:00 234.4 F 8.2
12:00 240.0 F 10.0
14:00 249.2 F 14.7
16:00 261.8 F 21.8
18:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.7
2:00 271.4 F 28.1
3:00 272.5 F 28.8
4:00 271.6 F 28.2

Min 271.3 F 28.0
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 220.2 F 0.0
10:00 236.7 F 0.0
15:00 235.1 F 0.0
20:00 227.6 F 0.0
25:00 225.8 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 18:20
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

05/30/2023 04:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____