

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3652

Name: RITTER2  
Sterilizer ID: V2313228  
05/28/2024 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.9 F 0.0  
2:00 137.0 F 1.5  
4:00 146.5 F 4.3  
6:00 173.5 F 7.6  
8:00 193.4 F 11.4  
10:00 209.8 F 4.3  
12:00 216.8 F 8.5  
14:00 227.8 F 12.6  
16:00 231.2 F 9.0  
18:00 239.2 F 11.5  
20:00 243.0 F 12.7  
22:00 254.1 F 18.4  
24:00 264.1 F 24.5  
26:00 272.1 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.5  
1:00 271.6 F 29.9  
2:00 272.6 F 30.0  
3:00 271.3 F 29.4  
4:00 271.6 F 30.0

Min 271.1 F 29.3  
Max 272.8 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 202.5 F 0.0  
10:00 206.8 F 0.0  
15:00 201.9 F 0.0  
20:00 193.7 F 0.0  
25:00 191.7 F 0.0  
30:00 191.1 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 26:05  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:04:52  
CYCLE COMPLETE

05/28/2024 01:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_