
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001299
11:34 14/10/2022
WRAPPED
135 C for 10 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 5:01
STERILIZING 6:02
136.4 C 224kPa 6:02
136.3 C 224kPa 6:32
136.4 C 224kPa 7:02
136.4 C 224kPa 7:32
136.3 C 224kPa 8:02
136.4 C 223kPa 8:32
136.4 C 225kPa 9:02
136.4 C 225kPa 9:32
136.1 C 222kPa 10:02
136.3 C 224kPa 10:32
136.2 C 224kPa 11:02
136.3 C 225kPa 11:32
136.2 C 222kPa 12:02
136.2 C 223kPa 12:32
136.4 C 224kPa 13:02
136.4 C 224kPa 13:32
136.4 C 225kPa 14:02
136.2 C 223kPa 14:32
136.3 C 224kPa 15:02
136.1 C 222kPa 15:32
136.4 C 225kPa 16:02
136.4 C 225kPa 16:02
VENTING 16:02
AIR DRYING 16:55
CYCLE COMPLETE 31:55

Digital Signature #
7300C7B95C71D9EC
