
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000143
8:28 14/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:26
PRESSURIZING 4:27
STERILIZING 5:55
136.0 C 220kPa 5:55
136.4 C 225kPa 6:25
136.4 C 225kPa 6:55
136.3 C 225kPa 7:25
136.4 C 225kPa 7:55
136.4 C 224kPa 8:25
136.3 C 223kPa 8:55
136.4 C 225kPa 9:25
136.4 C 225kPa 9:25
VENTING 9:25
AIR DRYING 10:33
CYCLE COMPLETE 12:55

Digital Signature #
200A99999B4DAD17
