

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5288

Name: RITTER 1  
Sterilizer ID: V2178198  
01/28/2025 04:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.0 F 0.0  
2:00 169.1 F 3.6  
4:00 179.5 F 8.1  
6:00 202.4 F 12.8  
8:00 213.7 F 6.3  
10:00 226.4 F 11.2  
12:00 228.8 F 7.3  
14:00 242.3 F 12.9  
16:00 240.8 F 11.3  
18:00 253.1 F 17.4  
20:00 264.1 F 23.9  
22:00 271.8 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.1 F 29.6  
2:00 272.6 F 29.5  
3:00 272.2 F 29.2  
4:00 271.4 F 28.8

Min 271.2 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 215.6 F 0.0  
10:00 214.7 F 0.0  
15:00 214.4 F 0.0  
20:00 216.2 F 0.0  
25:00 223.1 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 22:09  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:01:06  
CYCLE COMPLETE

01/28/2025 05:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_