

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4374

Name: RITTER 1  
Sterilizer ID: V2178198  
03/11/2024 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.1 F 0.0  
2:00 71.2 F 0.2  
4:00 95.2 F 1.1  
6:00 116.2 F 3.3  
8:00 140.4 F 6.2  
10:00 168.2 F 9.6  
12:00 186.1 F 12.9  
14:00 203.4 F 3.6  
16:00 213.4 F 6.6  
18:00 223.2 F 9.9  
20:00 230.9 F 11.3  
22:00 233.5 F 9.4  
24:00 240.1 F 11.0  
26:00 242.6 F 11.9  
28:00 253.8 F 17.6  
30:00 263.9 F 23.6  
32:00 272.3 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.5 F 29.2  
2:00 271.6 F 28.7  
3:00 272.4 F 29.1  
4:00 271.5 F 28.7  
  
Min 271.5 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
5:00 210.6 F 0.0  
10:00 217.5 F 0.0  
15:00 216.8 F 0.0  
20:00 222.1 F 0.0  
25:00 225.8 F 0.0  
30:00 224.2 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 32:14  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:10:32  
CYCLE COMPLETE

03/11/2024 10:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_