
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001012
14:20 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:05
STERILIZING 3:11
136.4 C 233kPa 3:11
136.2 C 232kPa 3:41
136.3 C 233kPa 4:11
136.2 C 231kPa 4:41
136.2 C 233kPa 5:11
136.0 C 231kPa 5:41
136.4 C 232kPa 6:11
136.4 C 233kPa 6:41
136.4 C 233kPa 6:41
VENTING 6:41
AIR DRYING 7:34
CYCLE COMPLETE 7:37

Digital Signature #
FA46230CAC9CABA4
