
Midmark M11 - v1.0.5
Total Cycles: 1086

Name: RITTER2
Sterilizer ID: V2313228
05/10/2022 07:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 73.8 F 0.0
4:00 85.4 F 0.0
6:00 107.5 F 0.0
8:00 154.5 F 0.0
10:00 179.6 F 1.7
12:00 193.3 F 5.3
14:00 207.3 F 8.9
16:00 218.7 F 12.6
18:00 226.0 F 7.0
20:00 236.5 F 11.5
22:00 234.7 F 8.6
24:00 242.3 F 11.3
26:00 245.6 F 13.1
28:00 257.0 F 19.2
30:00 267.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.0 F 29.1
2:00 272.2 F 28.8
3:00 271.7 F 28.9
4:00 272.5 F 29.1

Min 271.4 F 28.3
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 202.8 F 0.0
10:00 206.4 F 0.0
15:00 206.7 F 0.0
20:00 205.1 F 0.0
25:00 204.0 F 0.0
30:00 204.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 31:24
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:19
CYCLE COMPLETE

05/10/2022 08:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____