

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4381

Name: RITTER2  
Sterilizer ID: V2313228  
02/26/2025 11:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.0 F 0.0  
2:00 161.1 F 2.0  
4:00 162.7 F 5.4  
6:00 192.3 F 9.6  
8:00 211.4 F 8.5  
10:00 219.4 F 7.6  
12:00 230.7 F 12.1  
14:00 233.5 F 9.0  
16:00 240.8 F 10.9  
18:00 245.3 F 13.4  
20:00 256.8 F 19.6  
22:00 267.0 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.4 F 29.9  
2:00 271.9 F 29.3  
3:00 271.2 F 29.4  
4:00 272.4 F 29.9

Min 271.0 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.3 F 0.0  
5:00 209.0 F 0.0  
10:00 204.9 F 0.0  
15:00 193.0 F 0.0  
20:00 195.0 F 0.0  
25:00 188.3 F 0.0  
30:00 182.5 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 23:36  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:03:30  
CYCLE COMPLETE

02/26/2025 12:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_