
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001333
15:29 03/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:55
136.4 C 220kPa 3:55
136.4 C 225kPa 4:25
136.1 C 223kPa 4:55
136.4 C 225kPa 5:25
136.3 C 224kPa 5:55
136.3 C 225kPa 6:25
136.1 C 223kPa 6:55
136.2 C 223kPa 7:25
136.2 C 223kPa 7:25
VENTING 7:25
AIR DRYING 8:14
CYCLE COMPLETE 23:14

Digital Signature #
98C268FDDF3CC34D
