
Midmark M11 - v1.0.5
Total Cycles: 152

Name: RITTER2
Sterilizer ID: V2313228
06/29/2021 08:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.6 F 0.0
2:00 73.7 F 0.0
4:00 77.3 F 1.0
6:00 97.4 F 3.1
8:00 130.8 F 6.2
10:00 164.1 F 10.0
12:00 191.3 F 6.3
14:00 203.1 F 6.8
16:00 217.6 F 11.0
18:00 227.0 F 6.8
20:00 233.6 F 12.5
22:00 236.3 F 10.4
24:00 247.6 F 16.6
26:00 260.4 F 23.0
28:00 270.4 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.9
1:00 272.4 F 31.3
2:00 272.8 F 31.1
3:00 272.7 F 30.8
4:00 272.3 F 30.9

Min 272.1 F 30.8
Max 273.1 F 31.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.0 F 0.0
5:00 231.5 F 0.0
10:00 215.9 F 0.0
15:00 224.5 F 0.0
20:00 198.4 F 0.0
25:00 217.1 F 0.0
30:00 211.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:43
STERILIZING: 4:00
VENTING: 2:54
DRYING: 30:00
TOTAL CYCLE: 01:06:34
CYCLE COMPLETE

06/29/2021 09:57 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____