

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2821

Name: RITTER 1  
Sterilizer ID: V2178198  
01/25/2023 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.8 F 0.0  
2:00 180.8 F 0.0  
4:00 195.9 F 4.7  
6:00 217.6 F 9.8  
8:00 224.9 F 6.3  
10:00 236.5 F 11.1  
12:00 235.0 F 8.6  
14:00 239.8 F 9.9  
16:00 248.3 F 14.3  
18:00 260.9 F 21.3  
20:00 272.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.2  
1:00 272.6 F 29.0  
2:00 271.8 F 28.2  
3:00 272.6 F 29.0  
4:00 272.1 F 28.5

Min 271.5 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 221.6 F 0.0  
10:00 234.0 F 0.0  
15:00 234.0 F 0.0  
20:00 228.0 F 0.0  
25:00 225.7 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 1:21  
HEATING: 20:23  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 00:58:59  
CYCLE COMPLETE

01/25/2023 03:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_