

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3993

Name: RITTER2  
Sterilizer ID: V2313228  
10/03/2024 10:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.0 F 0.0  
2:00 155.6 F 1.6  
4:00 164.9 F 4.6  
6:00 184.5 F 8.3  
8:00 210.1 F 12.3  
10:00 217.2 F 6.6  
12:00 228.8 F 10.9  
14:00 229.2 F 7.6  
16:00 238.8 F 12.7  
18:00 239.4 F 11.1  
20:00 251.3 F 17.0  
22:00 261.7 F 23.1  
24:00 271.3 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 30.0  
1:00 272.6 F 30.3  
2:00 271.9 F 30.4  
3:00 272.0 F 30.1  
4:00 271.9 F 30.5

Min 271.4 F 29.9  
Max 273.1 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.9 F 0.0  
5:00 211.5 F 0.0  
10:00 206.7 F 0.0  
15:00 199.7 F 0.0  
20:00 191.6 F 0.0  
25:00 189.5 F 0.0  
30:00 188.7 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 24:38  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:03:43  
CYCLE COMPLETE

10/03/2024 11:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_