
Midmark M11 - v1.0.5
Total Cycles: 4080

Name: RITTER2
Sterilizer ID: V2313228
11/04/2024 12:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.2 F 0.0
2:00 166.5 F 2.3
4:00 194.2 F 6.8
6:00 210.3 F 11.9
8:00 220.7 F 7.6
10:00 234.1 F 13.0
12:00 234.8 F 10.5
14:00 236.7 F 8.8
16:00 248.7 F 15.7
18:00 260.4 F 22.3
20:00 271.3 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 273.1 F 30.3
2:00 271.1 F 29.5
3:00 273.0 F 30.4
4:00 273.0 F 30.5

Min 271.0 F 29.5
Max 273.2 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.5 F 0.0
5:00 231.3 F 0.0
10:00 228.2 F 0.0
15:00 229.0 F 0.0
20:00 224.6 F 0.0
25:00 211.4 F 0.0
30:00 212.8 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

11/04/2024 01:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____