

-----  
Midmark M11 - v1.0.0  
Total Cycles: 656

Name: MIDMARK  
Sterilizer ID: V2178198  
02/09/2021 10:17 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 103.8 F 0.0  
2:00 104.1 F 0.0  
4:00 167.2 F 0.2  
6:00 148.1 F 2.6  
8:00 172.1 F 5.8  
10:00 211.6 F 10.2  
12:00 228.6 F 6.3  
14:00 223.7 F 8.2  
16:00 235.3 F 12.8  
18:00 236.9 F 9.5  
20:00 238.9 F 9.2  
22:00 247.1 F 13.4  
24:00 258.6 F 19.6  
26:00 269.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.7 F 28.8  
2:00 272.5 F 28.8  
3:00 272.2 F 28.9  
4:00 271.7 F 28.7

Min 270.8 F 27.8  
Max 273.0 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
5:00 223.6 F 0.0  
10:00 229.9 F 0.0  
15:00 237.5 F 0.0  
20:00 216.1 F 0.0  
25:00 196.2 F 0.0  
30:00 184.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:57  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:04:60  
CYCLE COMPLETE

02/09/2021 11:22 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01