
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007898
13:46 26/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:02
STERILIZING 3:02
136.5 C 220kPa 3:02
136.3 C 224kPa 3:32
136.3 C 220kPa 4:02
136.5 C 225kPa 4:32
136.4 C 224kPa 5:02
136.4 C 222kPa 5:32
136.4 C 224kPa 6:02
136.4 C 224kPa 6:32
136.4 C 224kPa 6:32
VENTING 6:32
AIR DRYING 7:24
CYCLE COMPLETE 24:02

Digital Signature #
2331392DA1C1CD5B
