
Midmark M11 - v1.0.0
Total Cycles: 3977

Name: RITTER 1
Sterilizer ID: V2178198
11/03/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.6 F 0.0
2:00 167.7 F 1.8
4:00 166.3 F 4.7
6:00 186.8 F 8.8
8:00 209.0 F 12.0
10:00 218.9 F 6.7
12:00 230.7 F 11.3
14:00 230.3 F 7.3
16:00 242.6 F 12.9
18:00 241.5 F 11.1
20:00 253.7 F 17.3
22:00 264.7 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.0
2:00 271.5 F 28.3
3:00 272.3 F 29.0
4:00 272.4 F 28.8

Min 271.3 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 215.8 F 0.0
10:00 220.2 F 0.0
15:00 222.6 F 0.0
20:00 228.6 F 0.0
25:00 229.5 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:59
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:02:50
CYCLE COMPLETE

11/03/2023 10:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____