
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002481
13:09 06/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:25
STERILIZING 3:24
136.2 C 220kPa 3:24
136.4 C 226kPa 3:54
136.4 C 224kPa 4:24
136.4 C 225kPa 4:54
136.1 C 223kPa 5:24
136.1 C 223kPa 5:54
136.4 C 225kPa 6:24
136.3 C 225kPa 6:54
136.3 C 225kPa 6:54
VENTING 6:54
AIR DRYING 7:47
CYCLE COMPLETE 37:47

Digital Signature #
FA6E7CFEE119669E
