
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007420
9:30 24/05/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:27
STERILIZING 3:33
136.4 C 220kPa 3:33
136.1 C 223kPa 4:03
136.4 C 224kPa 4:33
136.4 C 223kPa 5:03
136.4 C 224kPa 5:33
136.4 C 224kPa 6:03
136.4 C 225kPa 6:33
136.3 C 220kPa 7:03
136.3 C 220kPa 7:03
VENTING 7:03
AIR DRYING 7:51
CYCLE COMPLETE 32:00

Digital Signature #
DD0E8D116ED253F7
