
Midmark M11 - v1.0.0
Total Cycles: 3950

Name: RITTER 1
Sterilizer ID: V2178198
10/27/2023 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.7 F 0.0
2:00 162.8 F 2.3
4:00 165.0 F 5.8
6:00 190.4 F 10.1
8:00 219.3 F 5.1
10:00 225.5 F 8.1
12:00 236.8 F 12.8
14:00 235.6 F 9.3
16:00 240.5 F 9.8
18:00 245.9 F 13.2
20:00 257.8 F 19.6
22:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.9
2:00 271.3 F 28.3
3:00 272.5 F 29.0
4:00 271.8 F 28.4

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 232.6 F 0.0
10:00 223.5 F 0.0
15:00 220.5 F 0.0
20:00 215.9 F 0.0
25:00 223.6 F 0.0
30:00 233.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 23:11
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:01:43
CYCLE COMPLETE

10/27/2023 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____