

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 004030  
13:31 15/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:50  
PRESSURIZING 2:53  
STERILIZING 3:57  
136.0 C 221kPa 3:57  
136.4 C 223kPa 4:27  
136.3 C 223kPa 4:57  
136.4 C 224kPa 5:27  
136.1 C 223kPa 5:57  
136.4 C 224kPa 6:27  
136.1 C 221kPa 6:57  
136.4 C 225kPa 7:27  
136.4 C 225kPa 7:27  
VENTING 7:27  
AIR DRYING 8:19  
CYCLE COMPLETE 12:15

-----  
Digital Signature #  
975D8285C6364D81  
-----