
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002096
9:24 01/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:51
STERILIZING 4:02
136.3 C 223kPa 4:02
136.3 C 224kPa 4:32
136.4 C 225kPa 5:02
136.4 C 225kPa 5:32
136.4 C 223kPa 6:02
136.4 C 223kPa 6:32
136.2 C 224kPa 7:02
136.4 C 225kPa 7:32
136.4 C 225kPa 7:32
VENTING 7:32
AIR DRYING 8:24
CYCLE COMPLETE 22:41

Digital Signature #
A18043746DE3E48E
