
Midmark M11 - v1.0.0
Total Cycles: 1939

Name: RITTER 1
Sterilizer ID: V2178198
05/12/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.2 F 0.0
2:00 195.8 F 3.6
4:00 194.2 F 8.4
6:00 211.1 F 10.6
8:00 231.7 F 9.0
10:00 238.5 F 9.9
12:00 241.9 F 11.6
14:00 238.9 F 9.7
16:00 252.6 F 16.5
18:00 264.3 F 23.2
20:00 271.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.2 F 28.6
2:00 272.6 F 28.9
3:00 272.4 F 29.2
4:00 271.7 F 28.6

Min 271.7 F 28.5
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 223.8 F 0.0
10:00 233.7 F 0.0
15:00 228.9 F 0.0
20:00 226.0 F 0.0
25:00 224.0 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:24
CYCLE COMPLETE

05/12/2022 12:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____