
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 002334
12:31 29/11/2024
AIR DRYING ONLY
60 min. MAXIMUM
CYCLE START 0:00
AIR DRYING 0:11
CYCLE COMPLETE
1:00:12

Digital Signature #
ACBDB9FC0694C024
