
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000105
14:26 02/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 3:45
STERILIZING 5:18
136.0 C 222kPa 5:18
136.4 C 226kPa 5:48
136.4 C 223kPa 6:18
136.4 C 227kPa 6:48
136.4 C 226kPa 7:18
136.2 C 224kPa 7:48
136.1 C 224kPa 8:18
136.0 C 224kPa 8:48
136.0 C 224kPa 8:48
VENTING 8:48
AIR DRYING 9:52
CYCLE COMPLETE 24:09

Digital Signature #
13B53F03BF6CE6D8
