
Midmark M11 - v1.0.5
Total Cycles: 1179

Name: RITTER2
Sterilizer ID: V2313228
06/09/2022 08:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 146.7 F 1.1
4:00 153.4 F 4.2
6:00 178.2 F 8.6
8:00 202.4 F 12.0
10:00 220.6 F 7.2
12:00 231.0 F 12.1
14:00 233.5 F 9.2
16:00 238.9 F 10.1
18:00 245.3 F 13.8
20:00 257.2 F 20.1
22:00 267.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.4
1:00 271.7 F 29.9
2:00 272.5 F 30.4
3:00 272.8 F 30.2
4:00 272.5 F 30.0

Min 271.1 F 29.2
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.1 F 0.0
5:00 217.5 F 0.0
10:00 225.5 F 0.0
15:00 225.9 F 0.0
20:00 222.7 F 0.0
25:00 221.8 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:32
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:01:41
CYCLE COMPLETE

06/09/2022 09:24 AM
ACCEPTED: X
REJECTED:
Operator: U02