

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2458

Name: RITTER2  
Sterilizer ID: V2313228  
06/02/2023 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 118.0 F 0.0  
2:00 119.8 F 0.0  
4:00 127.0 F 0.3  
6:00 158.8 F 3.2  
8:00 184.4 F 6.5  
10:00 203.1 F 10.4  
12:00 217.1 F 6.6  
14:00 226.3 F 9.1  
16:00 235.9 F 11.7  
18:00 237.9 F 11.0  
20:00 236.1 F 9.2  
22:00 249.2 F 15.4  
24:00 260.1 F 21.5  
26:00 270.0 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.7 F 29.8  
2:00 271.7 F 29.1  
3:00 272.2 F 29.9  
4:00 272.3 F 29.5

Min 271.3 F 29.0  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
5:00 205.3 F 0.0  
10:00 221.6 F 0.0  
15:00 217.6 F 0.0  
20:00 216.1 F 0.0  
25:00 217.4 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 26:58  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:05:19  
CYCLE COMPLETE

06/02/2023 01:08 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_