
Midmark M11 - v1.0.0
Total Cycles: 5218

Name: RITTER 1
Sterilizer ID: V2178198
01/02/2025 03:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.2 F 0.0
2:00 136.5 F 0.1
4:00 161.7 F 0.6
6:00 186.2 F 3.6
8:00 203.5 F 7.1
10:00 215.5 F 11.2
12:00 221.4 F 5.1
14:00 232.9 F 10.0
16:00 235.5 F 9.0
18:00 243.5 F 12.5
20:00 241.5 F 11.0
22:00 253.8 F 17.3
24:00 264.8 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.1
2:00 271.9 F 28.5
3:00 271.5 F 28.4
4:00 272.6 F 29.1

Min 271.4 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 206.3 F 0.0
10:00 207.5 F 0.0
15:00 202.3 F 0.0
20:00 206.5 F 0.0
25:00 205.4 F 0.0
30:00 202.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:59
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:38
CYCLE COMPLETE

01/02/2025 04:22 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____