

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001438  
9:41 18/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:32  
PRESSURIZING 3:34  
STERILIZING 4:41  
136.4 C 223kPa 4:41  
136.3 C 224kPa 5:11  
136.4 C 224kPa 5:41  
136.4 C 224kPa 6:11  
136.4 C 224kPa 6:41  
136.4 C 223kPa 7:11  
136.4 C 224kPa 7:41  
136.4 C 224kPa 8:11  
136.4 C 224kPa 8:11  
VENTING 8:11  
AIR DRYING 9:07  
CYCLE COMPLETE 16:30  
-----

Digital Signature #  
886F3EE99D1BB8D6  
-----