
Midmark M11 - v1.0.5
Total Cycles: 1958

Name: RITTER2
Sterilizer ID: V2313228
01/23/2023 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.2 F 0.0
2:00 179.3 F 0.2
4:00 192.9 F 2.2
6:00 209.9 F 6.7
8:00 227.0 F 11.5
10:00 230.8 F 8.3
12:00 239.7 F 11.6
14:00 243.9 F 13.0
16:00 244.1 F 12.7
18:00 257.0 F 19.5
20:00 268.0 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.4 F 29.7
2:00 272.3 F 29.3
3:00 271.5 F 28.8
4:00 271.3 F 29.1

Min 271.2 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 229.2 F 0.0
10:00 222.2 F 0.0
15:00 219.9 F 0.0
20:00 222.2 F 0.0
25:00 221.4 F 0.0
30:00 223.7 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

01/23/2023 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____