
Midmark M11 - v1.0.5
Total Cycles: 3362

Name: RITTER2
Sterilizer ID: V2313228
02/07/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.1 F 0.0
2:00 81.7 F 0.0
4:00 92.1 F 0.0
6:00 96.6 F 0.9
8:00 137.5 F 3.5
10:00 166.5 F 6.3
12:00 183.9 F 9.7
14:00 198.5 F 10.3
16:00 220.7 F 5.6
18:00 229.3 F 9.2
20:00 240.4 F 13.3
22:00 236.0 F 9.4
24:00 239.5 F 10.4
26:00 245.1 F 12.9
28:00 256.6 F 19.0
30:00 266.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.7 F 29.2
2:00 271.1 F 28.7
3:00 272.4 F 29.2
4:00 272.2 F 29.1

Min 271.0 F 28.5
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 229.9 F 0.0
10:00 216.0 F 0.0
15:00 187.8 F 0.0
20:00 174.1 F 0.0
25:00 170.5 F 0.0
30:00 168.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:35
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:45
CYCLE COMPLETE

02/07/2024 11:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____