

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1934

Name: RITTER 1  
Sterilizer ID: V2178198  
05/11/2022 03:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.3 F 0.0  
2:00 189.5 F 4.0  
4:00 201.4 F 8.9  
6:00 218.6 F 8.1  
8:00 229.0 F 9.4  
10:00 236.1 F 9.0  
12:00 243.4 F 12.6  
14:00 242.2 F 11.3  
16:00 255.5 F 18.1  
18:00 267.0 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.3 F 28.6  
2:00 271.7 F 28.4  
3:00 272.3 F 28.5  
4:00 271.6 F 28.4  
  
Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.2 F 0.0  
5:00 236.9 F 0.0  
10:00 231.0 F 0.0  
15:00 230.3 F 0.0  
20:00 230.9 F 0.0  
25:00 226.3 F 0.0  
30:00 220.5 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:30  
CYCLE COMPLETE

05/11/2022 04:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_