
Midmark M11 - v1.0.0
Total Cycles: 3000

Name: RITTER 1
Sterilizer ID: V2178198
03/14/2023 08:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.1 F 0.0
2:00 87.3 F 0.0
4:00 94.1 F 0.0
6:00 125.1 F 0.5
8:00 156.5 F 2.4
10:00 180.7 F 4.9
12:00 198.4 F 8.1
14:00 211.8 F 11.6
16:00 217.3 F 4.6
18:00 228.5 F 8.6
20:00 238.9 F 12.9
22:00 237.1 F 9.7
24:00 237.9 F 9.2
26:00 246.7 F 13.6
28:00 258.0 F 19.6
30:00 268.3 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.8
2:00 271.6 F 28.3
3:00 272.5 F 28.8
4:00 271.5 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 205.5 F 0.0
10:00 224.0 F 0.0
15:00 228.7 F 0.0
20:00 230.4 F 0.0
25:00 233.5 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:17
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:18
CYCLE COMPLETE

03/14/2023 09:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____