

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2422

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2023 01:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.9 F 0.0  
2:00 180.2 F 3.4  
4:00 189.5 F 8.1  
6:00 214.6 F 10.4  
8:00 227.4 F 9.2  
10:00 234.0 F 9.5  
12:00 241.6 F 12.7  
14:00 241.7 F 11.6  
16:00 255.0 F 18.5  
18:00 266.4 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.6 F 29.7  
2:00 271.5 F 29.0  
3:00 272.7 F 29.8  
4:00 271.9 F 29.2  
  
Min 271.5 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.9 F 0.0  
5:00 213.4 F 0.0  
10:00 222.2 F 0.0  
15:00 224.9 F 0.0  
20:00 225.4 F 0.0  
25:00 224.4 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 19:34  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

05/26/2023 02:54 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_