
Midmark M11 - v1.0.0
Total Cycles: 5467

Name: RITTER 1
Sterilizer ID: V2178198
04/02/2025 08:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 69.7 F 0.4
4:00 90.6 F 2.0
6:00 130.4 F 4.5
8:00 156.1 F 7.5
10:00 176.5 F 11.1
12:00 189.4 F 2.3
14:00 210.8 F 5.1
16:00 221.1 F 8.6
18:00 231.5 F 12.5
20:00 230.8 F 7.6
22:00 242.6 F 12.6
24:00 239.9 F 10.3
26:00 252.4 F 16.4
28:00 263.0 F 22.6
30:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.6 F 28.7
2:00 272.1 F 28.6
3:00 272.4 F 28.5
4:00 272.2 F 28.7

Min 271.6 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.0 F 0.0
5:00 182.5 F 0.0
10:00 190.6 F 0.0
15:00 197.5 F 0.0
20:00 202.1 F 0.0
25:00 203.1 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:17
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:08:28
CYCLE COMPLETE

04/02/2025 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____