
Midmark M11 - v1.0.0
Total Cycles: 5179

Name: RITTER 1
Sterilizer ID: V2178198
12/11/2024 12:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.9 F 0.0
2:00 101.8 F 0.0
4:00 118.7 F 0.0
6:00 152.5 F 0.1
8:00 178.6 F 0.3
10:00 199.8 F 2.9
12:00 219.2 F 6.5
14:00 225.9 F 10.4
16:00 231.3 F 8.0
18:00 237.8 F 11.2
20:00 235.3 F 8.8
22:00 240.3 F 10.3
24:00 247.4 F 13.9
26:00 259.0 F 20.2
28:00 269.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 271.7 F 28.7
2:00 272.2 F 29.1
3:00 272.6 F 29.4
4:00 272.8 F 29.3

Min 271.3 F 28.3
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.4 F 0.0
5:00 216.9 F 0.0
10:00 220.3 F 0.0
15:00 230.1 F 0.0
20:00 234.8 F 0.0
25:00 234.9 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 29:21
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:07:50
CYCLE COMPLETE

12/11/2024 01:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____