

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5359

Name: RITTER 1  
Sterilizer ID: V2178198  
02/24/2025 12:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.7 F 0.0  
2:00 170.7 F 2.8  
4:00 173.3 F 6.7  
6:00 203.2 F 10.6  
8:00 216.8 F 3.4  
10:00 223.2 F 7.8  
12:00 234.3 F 12.1  
14:00 233.7 F 8.4  
16:00 245.6 F 13.0  
18:00 243.2 F 12.0  
20:00 254.6 F 17.8  
22:00 264.9 F 24.0  
24:00 271.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.3 F 29.3  
2:00 272.2 F 28.9  
3:00 271.4 F 28.5  
4:00 271.4 F 28.8

Min 271.2 F 28.4  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 196.8 F 0.0  
10:00 195.1 F 0.0  
15:00 191.8 F 0.0  
20:00 194.8 F 0.0  
25:00 197.3 F 0.0  
30:00 199.4 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 24:05  
STERILIZING: 4:00  
VENTING: 3:58  
DRYING: 30:00  
TOTAL CYCLE: 01:03:05  
CYCLE COMPLETE

02/24/2025 01:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_