

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1897

Name: RITTER2  
Sterilizer ID: V2313228  
01/05/2023 08:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 79.2 F 0.0  
2:00 79.6 F 0.0  
4:00 87.1 F 0.0  
6:00 115.4 F 0.0  
8:00 158.3 F 0.1  
10:00 191.7 F 0.2  
12:00 204.9 F 3.5  
14:00 216.7 F 6.8  
16:00 226.4 F 10.5  
18:00 232.9 F 8.7  
20:00 237.8 F 11.1  
22:00 235.8 F 8.8  
24:00 240.6 F 10.6  
26:00 246.6 F 13.9  
28:00 258.3 F 20.3  
30:00 268.9 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.5 F 29.6  
2:00 271.8 F 28.8  
3:00 271.7 F 29.2  
4:00 272.5 F 29.4  
  
Min 271.3 F 28.6  
Max 272.6 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 211.8 F 0.0  
10:00 225.4 F 0.0  
15:00 230.2 F 0.0  
20:00 239.4 F 0.0  
25:00 220.2 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:11  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:09:05  
CYCLE COMPLETE

01/05/2023 09:49 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_