
Midmark M11 - v1.0.0
Total Cycles: 2453

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2022 03:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.9 F 0.0
2:00 188.1 F 3.5
4:00 194.9 F 8.2
6:00 215.4 F 9.4
8:00 229.9 F 9.2
10:00 234.3 F 9.6
12:00 242.3 F 12.5
14:00 241.1 F 11.1
16:00 254.6 F 17.9
18:00 266.2 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.3 F 29.3
2:00 271.8 F 28.6
3:00 271.5 F 28.9
4:00 272.5 F 29.3

Min 271.1 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.8 F 0.0
5:00 235.5 F 0.0
10:00 230.9 F 0.0
15:00 231.7 F 0.0
20:00 226.8 F 0.0
25:00 226.7 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

10/18/2022 04:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____