

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2795

Name: RITTER 1  
Sterilizer ID: V2178198  
01/19/2023 09:34 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
2:00 186.1 F 2.4  
4:00 190.1 F 6.5  
6:00 206.0 F 11.6  
8:00 220.4 F 5.4  
10:00 231.8 F 10.6  
12:00 231.2 F 7.2  
14:00 245.7 F 13.0  
16:00 244.0 F 12.2  
18:00 256.4 F 18.7  
20:00 267.6 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.6 F 29.1  
2:00 272.1 F 28.6  
3:00 271.2 F 28.2  
4:00 272.3 F 29.1  
  
Min 271.2 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 229.4 F 0.0  
10:00 229.4 F 0.0  
15:00 228.9 F 0.0  
20:00 231.3 F 0.0  
25:00 226.3 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 21:29  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:45  
CYCLE COMPLETE

01/19/2023 10:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_