
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008063
9:14 14/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:11
STERILIZING 3:05
136.0 C 220kPa 3:05
136.4 C 224kPa 3:35
136.4 C 222kPa 4:05
136.1 C 217kPa 4:35
136.4 C 223kPa 5:05
136.1 C 220kPa 5:35
136.4 C 219kPa 6:05
136.2 C 220kPa 6:35
136.2 C 220kPa 6:35
VENTING 6:35
AIR DRYING 7:25
CYCLE COMPLETE 16:47

Digital Signature #
0D37C83ECC491034
