
Midmark M11 - v1.0.0
Total Cycles: 5116

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2024 01:50 PM
C102: FILL MODE
STOP PRESSED
CHAMBER TEMP 133F
CHAMBER PRESS 0.0PSI

** ITEMS NOT STERILE **

11/20/2024 03:25 PM