

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1995

Name: RITTER 1  
Sterilizer ID: V2178198  
06/06/2022 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.8 F 0.0  
2:00 171.5 F 4.0  
4:00 185.8 F 8.3  
6:00 205.1 F 13.3  
8:00 221.0 F 7.4  
10:00 233.8 F 12.4  
12:00 234.0 F 9.1  
14:00 237.6 F 9.5  
16:00 246.9 F 13.8  
18:00 259.2 F 20.6  
20:00 270.6 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.6 F 29.1  
2:00 271.3 F 28.4  
3:00 272.6 F 29.2  
4:00 271.8 F 28.6  
  
Min 271.2 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 208.2 F 0.0  
10:00 215.7 F 0.0  
15:00 235.2 F 0.0  
20:00 220.2 F 0.0  
25:00 217.6 F 0.0  
30:00 214.7 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 20:49  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:11  
CYCLE COMPLETE

06/06/2022 02:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_