
Midmark M11 - v1.0.0
Total Cycles: 2612

Name: RITTER 1
Sterilizer ID: V2178198
11/28/2022 12:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.1 F 0.0
2:00 209.8 F 5.3
4:00 219.5 F 11.1
6:00 227.3 F 7.1
8:00 237.1 F 12.9
10:00 239.0 F 10.8
12:00 237.9 F 9.4
14:00 251.8 F 16.4
16:00 263.8 F 23.4
18:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.4 F 29.2
2:00 271.5 F 28.4
3:00 272.4 F 29.4
4:00 272.0 F 28.9

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.6 F 0.0
5:00 231.0 F 0.0
10:00 232.1 F 0.0
15:00 230.1 F 0.0
20:00 229.1 F 0.0
25:00 228.9 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 18:01
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

11/28/2022 01:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____