

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001396  
6:43 09/08/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:20  
PRESSURIZING 5:00  
STERILIZING 6:40  
136.0 C 220kPa 6:40  
136.4 C 223kPa 7:10  
136.1 C 220kPa 7:40  
136.3 C 223kPa 8:10  
136.2 C 221kPa 8:40  
136.0 C 220kPa 9:10  
136.4 C 222kPa 9:41  
136.4 C 223kPa 10:10  
VENTING 10:10  
AIR DRYING 11:07  
CYCLE COMPLETE 11:11  
-----

Digital Signature #  
8257A48F7D1A292E  
-----