
Midmark M11 - v1.0.0
Total Cycles: 1431

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2021 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.3 F 0.0
2:00 101.4 F 0.0
4:00 113.2 F 0.2
6:00 147.2 F 2.0
8:00 173.0 F 4.9
10:00 197.7 F 9.1
12:00 214.1 F 10.5
14:00 224.0 F 6.7
16:00 235.1 F 11.2
18:00 231.8 F 7.4
20:00 245.0 F 13.2
22:00 242.5 F 11.4
24:00 255.0 F 17.7
26:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.3 F 28.5
2:00 272.3 F 28.8
3:00 271.8 F 28.1
4:00 272.7 F 28.9

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.5 F 0.0
5:00 224.4 F 0.0
10:00 224.6 F 0.0
15:00 226.7 F 0.0
20:00 226.0 F 0.0
25:00 226.7 F 0.0
30:00 222.2 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 27:41
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:05:37
CYCLE COMPLETE

12/08/2021 11:21 AM
ACCEPTED: X
REJECTED:
Operator: U02